



HUMAN CARE

Your Life. Your Way.



Assembly Instructions / Monteringsanvisning

Slow Down Brake for Carl-Oskar & Rebel
Släpbröms till Carl-Oskar och Rebel

Art.nr 10624



Assembly Instructions

Art. nr: 10624

These assembly instructions explain how to mount the Slow Down Brake on the Carl-Oskar and Rebel rollators.

General information

The Slow Down Brake fits all widths and heights of the Carl-Oskar and Rebel rollators. It can also be mounted on Carl-Oskar RA and Walking Table, Rebel RA and Walking Table as well as on Carl-Oskar for heavy users and Rebel for heavy users.

Assembly Instructions

The Slow Down Brake is delivered in pairs, a brake for the right rear wheel and a brake for the left rear wheel. The knob on the side of the brake must be mounted on the outside of the rollator frame. (picture 1)



Unscrew the knob that holds it together the clamp. (picture 2)



Insert the loose cassette with the smaller wheel at the bottom of the Slow Down Brake. It will be loose until the brake is mounted on the rollator. The upper adjustment screws must be fully unscrewed before the Slow Down Brake is fitted (Fig. 3)



Open the clamp and ensure that the smaller cassette remains in the Slow Down Brake and that the smaller wheel in the cassette rests against the rollator's rear tire. (picture 4)



The distance between the lower edge of the Slow Down Brake and the upper part of the rear fork should be approximately 2 cm. (picture 5)



Tighten the knob on the side so that the brake is firmly attached to the frame and the small wheel rests against the rear tire, to get the desired resistance on the wheel. Adjust the hardness with the upper knob. Lock the upper knob with the wing nut. Test the rollator with the Slow Down Brake to ensure that the resistance is as desired. (picture 6)



Repeat this process on the other side.

Monteringsanvisning

Art. nr: 10624

Den här monteringsanvisningen förklarar hur släpbromsen ska monteras på rollatorerna Carl-Oskar och Rebel.

Allmän information

Släpbromsen passar på alla bredder och höjder av rollatorerna Carl-Oskar och Rebel. Den kan även monteras på Carl-Oskar RA och Gåbord, Rebel RA och Gåbord samt på Carl-Oskar för tunga brukare och Rebel för tunga brukaren.

Monteringsanvisning

Släpbromsen levereras i par, en broms för höger bakhjul och en broms för vänster bakhjul. Vredet på sidan av släpbromsen skall vara monterat på utsidan av rollatorns ram. (bild 1)



Skruva loss vredet som håller ihop klämman. (bild 2)



För in den lösa kassetten med det mindre hjulet i botten på släpbromsen. Den kommer vara rörlig fram tills att bromsen monteras på rollatorn. De övre justeringsskruvarna skall vara helt uppskruvade innan släpbromsen monteras. (bild 3)



Öppna klamman och säkerställ att den mindre kassetten är kvar i släpbromsen och att det mindre hjulet i kassetten ligger mot rollatorns bakdäck. (bild 4)



Avståndet mellan nederkanten på släpbromsen och överdelen på bakgaffeln skall vara cirka 2 cm. (bild 5)



Dra åt vredet på sidan så att bromsen sitter stadigt fast på ramen och det lilla hjulet ligger an mot bakdäcket, för att få det önskade trögheten på hjulet. Justera hårdheten med det övre vredet. Lås den övre ratten med vingmuttern. Prova att rollatorn rullar som önskas. (bild 6)



Upprepa denna process på det den andra sidan.

HUMAN CARE SWEDEN (HQ)

Årstaängsvägen 21B

117 43 Stockholm

Phone:- +46 8 665 35 00

Fax:- +46 8 665 35 10

Email: info.se@humancaregroup.com

HUMAN CARE CANADA

10-155 Colonnade Road

Ottawa, ON K2E 7K1

Phone:- 613.723.6734

Fax:- 613.723.1058

Email: info.ca@humancaregroup.com

HUMAN CARE UNITED STATES

8006 Cameron Street, Suite K

Austin, TX 78754 USA

Phone: 512.476.7199

Fax: 512.476.7190

Email: info.us@humancaregroup.com

HUMAN CARE NEDERLAND

Elspeterweg 124

8076 PA Vierhouten

Phone: +31 577 412 171

Fax: +31 577 412 170

Email: info.nl@humancaregroup.com

www.humancaregroup.com

